



Contribution ID: 2026 Contribution code: WEPA110

Type: Poster Presentation

166.6 MHz HOM damped copper cavity designed for 4th generation synchrotron radiation sources

Wednesday, 10 May 2023 16:30 (2 hours)

The fourth-generation synchrotron light sources aim to achieve ultralow emittances and have very small dynamic aperture, which are expected to adopt on-axis injection schemes. Lower frequency rf systems are required for a large separation between RF buckets due to the limitations of kicker. We designed a 166.6 MHz normal conducting cavity with HOM damped used for the main cavity of storage ring. In this paper, we present the simulation studies of cavity including electromagnetic, HOMs, mechanical, and thermal calculations. A compact beam line absorber, which is uncommon in the copper cavities, is adopted to damp the harmful HOMs. The simulation results show that it can deeply damp the HOMs, but has no effect on the accelerating performance.

Funding Agency

Guangdong Basic and Applied Basic Research Foundation
(No. 2019B1515120046)

Footnotes

I have read and accept the Privacy Policy Statement

Yes

Primary authors: WU, Bin (Institute of High Energy Physics); ZHANG, Chunlin (Chinese Academy of Sciences); YU, Jiebing (Institute of High Energy Physics); LIU, Shenghua (Institute of High Energy Physics); LONG, Wei (Chinese Academy of Sciences); LI, Xiao (Institute of High Energy Physics); LIU, Yang (Dongguan Neutron Science Center); LU, Zhijun (Institute of High Energy Physics); ZHU, Junyu (Chinese Academy of Sciences); CHEN, Shengyi (Dongguan Neutron Science Center)

Presenter: ZHU, Junyu (Chinese Academy of Sciences)

Session Classification: Wednesday Poster Session

Track Classification: MC7: Accelerator Technology and Sustainability: MC7.T06: Room Temperature RF