IBIC2024 - 13th International Beam Instrumentation Conference



Contribution ID: 90 Contribution code: TUBC4

Type: Contributed Oral Presentation

Beam size measurement with gratings at BEPCII

Tuesday 10 September 2024 12:10 (20 minutes)

The vertical beam size measurement was carried out at BEPCII using a phase grating and an absorption grating based on the Talbot effect. Due to the partial coherence of the source, coherence length can be calculated by measuring the visibility decay of interferograms recorded at different distances behind gratings. The vertical beam size of 68.19±2µm was obtained based on the relationship between coherence length and source size. A comparison of the vertical emittance derived from grating Talbot method and synchrotron radiation visible light interferometer method was presented to evaluate the method. The vertical emittances from two methods are 1.41nm·rad and 1.40nm·rad respectively. The 0.1% difference indicates the grating Talbot method for beam size measurement is reliable. This technique has great potential in small beam size measurement in the fourth-generation synchrotron radiation light source, considering its small diffraction limitation and simple experimental setups.

Footnotes

Funding Agency

I have read and accept the Privacy Policy Statement

Yes

Primary author: ZHANG, Wan (Chinese Academy of Sciences)

Co-authors: Mr ZHU, Dechong (Institute of High Energy Physics); Mr CAO, Jianshe (Institute of High Energy Physics); HE, Jun (Institute of High Energy Physics); Mr YUE, Junhui (Institute of High Energy Physics); SUI, Yanfeng (Institute of High Energy Physics)

Presenter: ZHANG, Wan (Chinese Academy of Sciences)

Session Classification: TUB: Overview and Commissioning/Transverse Profile and Emittance Monitors

Track Classification: MC4: Transverse Profile and Emittance Monitors