

Contribution ID: 234 Contribution code: TUP04

Type: Poster Presentation

Development and performance evaluation of the cavity BPM system for SHINE

Tuesday 10 September 2024 16:00 (1h 30m)

The Shanghai high repetition rate XFEL and extreme light facility (SHINE) under construction is designed as one of the most advanced FEL facilities in the world, which will produce coherent x-rays with wavelengths from 0.05 to 3 nm and maximum repetition rate of 1MHz. To achieve precise beam trajectory measurement and stable alignment of the electron and photo beams in the undulator, the cavity beam position monitors (CBPM) including beam diameters of 35mm in LINAC and Bunch distribution section and 8mm in undulator have been designed and developed for the SHINE. The requirement of the transverse position resolution is better than 1µm and 200 nm for a single bunch of 100 pC, respectively. In this paper, we present the design of the cavity BPM system and the processing of the key equipment. The beam test bench has been established at the Shanghai Soft X-ray FEL facility (SXFEL), and preliminary beam experiments indicate that, with the bunch charge about 100pC, the position resolution of CBPM-35mm and CBPM-8mm is better than 330 nm and 70 nm, respectively.

Footnotes

Funding Agency

I have read and accept the Privacy Policy Statement

Yes

Primary author: CHEN, Jian (Shanghai Synchrotron Radiation Facility)

Co-authors: CHEN, Fangzhou (Shanghai Synchrotron Radiation Facility); LAI, Longwei (Shanghai Advanced

Research Institute); YUAN, Renxian (SSRF); CAO, Shanshan (Shanghai Advanced Research Institute)

Presenter: CHEN, Jian (Shanghai Synchrotron Radiation Facility)

Session Classification: TUP: Tuesday Poster Session

Track Classification: MC3: Beam Position Monitors