

Contribution ID: 251

Type: Poster Presentation

Design and deployment of an in-vacuum Electro-Optic BPM at the CERN SPS

Tuesday, 10 September 2024 16:00 (1h 30m)

Accurate monitoring and control of charged particle beams at the HL-LHC demands the development of new beam diagnostics tools. This poster provides an overview of the electro-optic beam position monitor (EO-BPM), currently taking measurements at CERN's SPS. This device uses the Pockels effect to monitor the transverse position and instabilities in the particle beam. Comprising of a laser source, electro-optic crystal, optical system, and a fast photodetector, the EO-BPM operates by generating a modulated optical signal directly linked to the propagating electric field of the beam.

The EO-BPM is designed as a self containing button with fibre-coupled laser connected to the crystal inside and a fibre coupled Mach-Zehnder interferometer yielding sum and difference signals on the outside. A segment of the SPS beam pipe is fitted with a mount to connect the button, allowing the electric field induced by the particle beam to be captured and transferred to the electro-optic crystal.

The goal is to gain insight into the transverse position along the bunch and the identification of intra-bunch instabilities, contributing to precision in beam monitoring and control.

Footnotes

Funding Agency

Funded by UKRI ST/N001583/1 and ST/P00203X/1, Royal Holloway University of London, CERN and EU Horizon 2020 GA No 101004730

I have read and accept the Privacy Policy Statement

Yes

Primary author: BOSMAN, Max (Royal Holloway, University of London)

Co-authors: GIBSON, Stephen (Royal Holloway, University of London); LEFEVRE, Thibaut (European Organi-

zation for Nuclear Research); LEVENS, Thomas (European Organization for Nuclear Research)

Presenter: BOSMAN, Max (Royal Holloway, University of London)

Session Classification: TUP: Tuesday Poster Session

Track Classification: MC3: Beam Position Monitors