



Contribution ID: **201** Contribution code: **WEP30**

Type: **Poster Presentation**

Detailed bench investigations and comparison of four low-light cameras

Wednesday 11 September 2024 14:20 (1h 30m)

Sensitive cameras are frequently operated to record low-light processes such as Beam Induced Fluorescence or optical transition radiation for transverse profile determination. We compared four cameras based on different principles: Firstly, we investigated an Image Intensifier equipped with a double MCP (producer ProxiVision); secondly, an electron-multiplied CCD (emCCD Teledyne Princeton Instruments ProEm+:512B); and, thirdly, sCMOS cameras (producer PCO.edge4.2bi and Teledyne Kinetix 22). LEDs generate light pulses within a wavelength range of 385 to 500 nm and a duration of 0.03 to 8 ms to vary the photon flux. Moreover, the spatial resolution is compared. The image intensifier is the most sensitive camera type and provides very low noise; however, the method provides only a limited spatial resolution. The investigated emCCD camera has a comparable sensitivity but provides a better spatial resolution. The sCMOS cameras provide a factor of about 5 to 10 lower sensitivity depending on wavelength. A quantitative comparison of signal-to-noise ratios and statical fluctuations for several wavelengths will be presented.

Footnotes

Funding Agency

I have read and accept the Privacy Policy Statement

Yes

Primary author: FORCK, Peter (GSI Helmholtzzentrum für Schwerionenforschung GmbH)

Co-authors: BAUER, Leonie (GSI Helmholtzzentrum für Schwerionenforschung GmbH); UDREA, Serban (GSI Helmholtzzentrum für Schwerionenforschung GmbH)

Presenter: FORCK, Peter (GSI Helmholtzzentrum für Schwerionenforschung GmbH)

Session Classification: WEP: Wednesday Poster Session

Track Classification: MC4: Transverse Profile and Emittance Monitors