



Contribution ID: 1750 Contribution code: TUPS036

Type: Poster Presentation

Design and EM simulations of 750 MHz IH-DTL tank for carbon ion in medical applications

Tuesday 3 June 2025 16:00 (2 hours)

This paper presents the design of 750 MHz IH-DTL (Interdigital H-mode Drift Tube Linac) tank, specifically developed to be part of a carbon ion injector for medical treatment applications. These sections provide a highly efficient solution for ion acceleration in the 5 to 10 MeV per nucleon energy range, offering a high shunt impedance. The study includes simulations of electromagnetic fields using CST Software, and beam dynamics simulations through a KONUS-type configuration

Footnotes

Paper preparation format

LaTeX

Region represented

Europe

Funding Agency

Author: MORENO, Gabriela (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas)

Co-authors: GAVELA, Daniel (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); GINER NAVARRO, Jorge (Instituto Universitario de Ciencia de los Materiales); RODRÍGUEZ PÁRAMO, Ángel (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); CALVO, Pedro (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); LOPEZ, Miguel (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); ETXEBARRIA, Jone (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); OLIVER, Concepcion (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); PEREZ MORALES, Jose (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas); LOPEZ, Michel (Added Value Solutions); Ms ALVARADO MARTIN, Maria (Added Value Solutions)

Presenter: MORENO, Gabriela (Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas)

Session Classification: Tuesday Poster Session

Track Classification: MC4: Hadron Accelerators: MC4.A08 Linear Accelerators