



Contribution ID: 656 Contribution code: TUPB050

Type: **Poster Presentation**

## Design study of HTS air-cored cyclotron coil system for medical RI production

*Tuesday 3 June 2025 16:00 (2 hours)*

The high current density of HTS material allows electromagnet to induce sufficiently strong magnetic field without relying on any iron core. This permits the design of air-cored cyclotron, where the absence of iron core brings the properties of light-weight and high field reproducibility, making it an ideal medical cyclotron to be installed inside hospitals. However, the cyclotron coil system need to induce highly accurate field while satisfying the engineering restriction from the HTS coil. Compact size, small fringe field and minimum fabrication cost are also desirable at the same time.

A HTS coil system of air-cored cyclotron is designed with the above restrictions taken into consideration. Multiple beam type accelerations that are required for medical RI production are simulated, in order to verify the usefulness of this design. In this work, the coil system design, the magnetic field and the HTS coil properties are presented. The feasibility of actual fabrication and in-hospital installation is discussed.

### Footnotes

### Paper preparation format

LaTeX

### Region represented

Asia

### Funding Agency

This work was supported by JSPS KAKENHI Grant Number JP24H00227, JP24H00227.

**Author:** CHONG, Tsun Him (Osaka University)

**Co-authors:** ISHIYAMA, Atshushi (Waseda University); KANDA, Hiroki (Osaka University); UEDA, Hiroshi (Osaka University); YOSHIDA, Jun (Sumitomo Heavy Industries Ltd.); FUKUDA, Mitsuhiro (Osaka University); NOGUCHI, So (Hokkaido University); YORITA, Tetsuhiko (Osaka University)

**Presenter:** CHONG, Tsun Him (Osaka University)

**Session Classification:** Tuesday Poster Session

**Track Classification:** MC8: Applications of Accelerators, and Engagement for Industry and Society: MC8.U04 Isotope Production