



Contribution ID: 452 Contribution code: MOPMO30

Type: **Poster Presentation**

## Development of a wide dynamic range and high-precision ammeter for beamline instrumentation at SIRIUS

*Monday 8 September 2025 16:00 (2 hours)*

Sirius beamlines require specialized electronic devices to monitor key parameters of the photon beam, such as position and flux, through the detection of extremely low-level electrical currents. Furthermore, experiments conducted in fly-scan mode usually demand fast, high-precision low-level current measurements. To address these requirements, the development of a wide dynamic range ammeter (from 1 pA to 10 mA) has been addressed, based on a logarithmic transimpedance amplifier, eliminating the need for scale switching and featuring high precision and fast response. The proposed device converts the input current into a logarithmic output voltage and offers two operation modes: the logarithmic ammeter and the logarithmic ratiometer, both of which are particularly useful in X-ray spectroscopy experiments. This paper considers an overview of the device as well as its preliminary characterization and results, including logarithmic conformity, bandwidth, and noise.

### Footnotes

### Funding Agency

The authors would like to thank the Brazilian Ministry of Science, Technology and Innovation for funding this project.

### I have read and accept the Conference Policies

Yes

**Author:** SILVA OLIVEIRA, Vinicius (Brazilian Synchrotron Light Laboratory)

**Co-authors:** CARDOSO, Fernando (Brazilian Synchrotron Light Laboratory); NALLIN, Patricia (Brazilian Synchrotron Light Laboratory)

**Presenter:** NALLIN, Patricia (Brazilian Synchrotron Light Laboratory)

**Session Classification:** MOP

**Track Classification:** MC07: Data Acquisition and Processing Platforms